

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204119	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 05/29/2013
LOCATION OF INSTRUMENT (STREET AND CITY) Shannon County Sheriff's Department		TIME OF INSPECTION 11:06 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 05/29/2013 11:06 A.M

☒ COMPUTER☒ DETECTOR☒ PROGRAM☒ FILTERS☒ HEATERS SAMPLE CHAMBER 49°C☒ QUARTZ STANDARD☒ FLOW DETECTOR☒ CALIBRATION☒ PUMP HIGH SPEED☒ PRINTER☒ INDICATOR LIGHTS

☒ SIMULATOR SOLUTION SUPPLIER Guth Laboratories, Inc. LOT # 12002 EXP. DATE 08/29/2014

☒ SIMULATOR TEMP (34°C ± 0.2°C) 33.98 °C SIMULATOR SN 1324 EXP. DATE 01/22/2014

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .099%

TEST 2 .100%

TEST 3 .101%

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
(USE OTHER SIDE IF NECESSARY).

This instrument is operating satisfactorily within the guidelines set by the Missouri Department of Health, Guth Laboratories Inc., 100% solution, Lot #12002, Bottle # 0354, expiration date 08/29/2014.

INSPECTING OFFICER

SIGNATURE

PRINT FULL NAME

Corporal Dale B. Pounds

TYPE II PERMIT NUMBER/EXPIRATION DATE

220325

09/28/2014

TELEPHONE NUMBER

(417) 469-3121

RETURN COMPLETED REPORT TO THE:

Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1209 gms/dl \pm .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 \pm 3% gms/210L Breath when heated to 34 Degrees Celsius \pm 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012
The expiration date for this lot number is August 29, 2014 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President
RepCo Marketing, Inc.

ACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204119
05/29/13

ARREST TIME: 10:30
SUBJECT NAME:
HOGAN/STEVEN L
DOB: 09/08/82 SEX: M
ARREST DATE: 09/28/14
ARRESTING OFFICER:
POUNDS/D/B
OFFICER I.D.: 984
ESTIMATING OFFICER:
SRME
OFFICER I.D.: SRME
ARREST NUMBER: 220325
ARREST DATE: 09/28/14
ISCELLANEOUS DATA:
--- BREATH ANALYSIS ---
AUDIO INTERFERENCE

DR SIGNATURE

Cpl D.B. Pounds

ACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204119
05/29/13

ARRESTING OFFICER:
POUNDS/D/B
OFFICER I.D.: 984
ARREST NUMBER: 220325
ARREST DATE: 09/28/14
ISCELLANEOUS DATA:
--- SUPERVISOR MODE ---

HK TEST	.000	11:09
ERNAL STANDARD	VERIFIED	11:09
ERNAL STANDARD	.099	11:10
HK TEST	.000	11:10
ERNAL STANDARD	.100	11:11
HK TEST	.000	11:11
ERNAL STANDARD	.101	11:12
HK TEST	.000	11:12

3
= .1
= .1

SIGNATURE

Cpl D.B. Pounds

ACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster

Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204119
05/29/13
11:06

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
ENTERS	
SAMPLE CHAMBER:	49C
LOW DETECTOR:	OKAY
UMP	
ISH SPEED:	OKAY
ECTOR:	OKAY
ILTERS:	OKAY
UARTZ STANDARD:	OKAY
ALIBRATION:	OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~

ATOR SIGNATURE

Cpl D.B. Pounds

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



DALE B POUNDS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 09/28/2012

Number 220325

Expires 09/28/2014

MO 580-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)